

Local Coverage Determination (LCD): Wheelchair Options/Accessories (L33792)

Links in PDF documents are not guaranteed to work. To follow a web link, please use the MCD Website.

Contractor Information

Contractor Name	Contract Type	Contract Number	Jurisdiction	State(s)
CGS Administrators, LLC	DME MAC	17013 -	DME MAC J-B	Illinois Indiana Kentucky Michigan Minnesota Ohio Wisconsin Alabama Arkansas Colorado Florida Georgia Louisiana Mississippi North Carolina New Mexico Oklahoma Puerto Rico South Carolina Tennessee Texas Virginia Virgin Islands West Virginia Connecticut District of Columbia Delaware Massachusetts Maryland Maine New Hampshire New Jersey New York - Entire State Pennsylvania Rhode Island Vermont Alaska American Samoa Arizona California - Entire State Guam Hawaii Iowa Idaho Kansas Missouri - Entire State Montana North Dakota Nebraska Nevada Oregon South Dakota
CGS Administrators, LLC	DME MAC	18003 -	DME MAC J-C	
Noridian Healthcare Solutions, LLC	DME MAC	16013 -	DME MAC J-A	
Noridian Healthcare Solutions, LLC	DME MAC	19003 -	DME MAC J-D	

LCD Information

Document Information

LCD ID
L33792Original Effective Date
For services performed on or after 10/01/2015Original ICD-9 LCD ID
[L27223](#)
[L11451](#)
[L11473](#)
[L11462](#)Revision Effective Date
For services performed on or after 01/01/2017Revision Ending Date
N/ALCD Title
Wheelchair Options/AccessoriesRetirement Date
N/A

AMA CPT / ADA CDT / AHA NUBC Copyright Statement
CPT only copyright 2002-2017 American Medical Association. All Rights Reserved. CPT is a registered trademark of the American Medical Association. Applicable FARS/DFARS Apply to Government Use. Fee schedules, relative value units, conversion factors and/or related components are not assigned by the AMA, are not part of CPT, and the AMA is not recommending their use. The AMA does not directly or indirectly practice medicine or dispense medical services. The AMA assumes no liability for data contained or not contained herein.

Notice Period Start Date
N/ANotice Period End Date
N/A

The Code on Dental Procedures and Nomenclature (Code) is published in Current Dental Terminology (CDT). Copyright © American Dental Association. All rights reserved. CDT and CDT-2016 are trademarks of the American Dental Association.

UB-04 Manual. OFFICIAL UB-04 DATA SPECIFICATIONS MANUAL, 2014, is copyrighted by American Hospital Association ("AHA"), Chicago, Illinois. No portion of OFFICIAL UB-04 MANUAL may be reproduced, sorted in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, recording or otherwise, without prior express, written consent of AHA." Health Forum reserves the right to change the copyright notice from time to time upon written notice to Company.

CMS National Coverage Policy Pub. 100-03 (Medicare National Coverage Determinations Manual), Chapter 1, Sections 280.1, 280.3

Coverage Guidance

Coverage Indications, Limitations, and/or Medical Necessity

For any item to be covered by Medicare, it must 1) be eligible for a defined Medicare benefit category, 2) be reasonable and necessary for the diagnosis or treatment of illness or injury or to improve the functioning of a

malformed body member, and 3) meet all other applicable Medicare statutory and regulatory requirements.

The purpose of a Local Coverage Determination (LCD) is to provide information regarding "reasonable and necessary" criteria based on Social Security Act § 1862(a)(1)(A) provisions.

In addition to the "reasonable and necessary" criteria contained in this LCD there are other payment rules, which are discussed in the following documents, that must also be met prior to Medicare reimbursement:

- The LCD-related Standard Documentation Requirements Article, located at the bottom of this policy under the Related Local Coverage Documents section.
- The LCD-related Policy Article, located at the bottom of this policy under the Related Local Coverage Documents section.
- Refer to the Supplier Manual for additional information on documentation requirements.
- Refer to the DME MAC web sites for additional bulletin articles and other publications related to this LCD.

For the items addressed in this LCD, the "reasonable and necessary" criteria, based on Social Security Act § 1862(a)(1)(A) provisions, are defined by the following coverage indications, limitations and/or medical necessity.

Options and accessories for wheelchairs are covered if the beneficiary has a wheelchair that meets Medicare coverage criteria and the option/accessory itself is medically necessary. Coverage criteria for specific items are described below.

If these criteria are not met, the item will be denied as not reasonable and necessary.

ARM OF CHAIR:

Adjustable arm height option (E0973, K0017, K0018, K0020) is covered if the beneficiary requires an arm height that is different from that available using nonadjustable arms and the beneficiary spends at least 2 hours per day in the wheelchair.

An arm trough (E2209) is covered if the beneficiary has quadriplegia, hemiplegia, or uncontrolled arm movements.

FOOTREST/ LEGREST:

Elevating legrests (E0990, K0046, K0047, K0053, K0195) are covered if:

1. The beneficiary has a musculoskeletal condition or the presence of a cast or brace which prevents 90 degree flexion at the knee; or
2. The beneficiary has significant edema of the lower extremities that requires an elevating legrest; or
3. The beneficiary meets the criteria for and has a reclining back on the wheelchair.

NONSTANDARD SEAT FRAME DIMENSIONS:

A nonstandard seat width and/or depth for a manual wheelchair (E2201-E2204) is covered only if the beneficiary's physical dimensions justify the need.

WHEELS/TIRES FOR MANUAL WHEELCHAIRS:

A gear reduction drive wheel (E2227) is covered if all of the following criteria are met:

1. The beneficiary has been self-propelling in a manual wheelchair for at least one year; and
2. The beneficiary has had a specialty evaluation that was performed by a licensed/certified medical professional, such as a PT or OT, or physician who has specific training and experience in rehabilitation wheelchair evaluations and that documents the need for the device in the beneficiary's home. The PT, OT, or physician may have no financial relationship with the supplier; and
3. The wheelchair is provided by a supplier that employs a RESNA-certified Assistive Technology Professional (ATP) who specializes in wheelchairs and who has direct, in-person involvement in the wheelchair selection for the beneficiary.

BATTERIES/ CHARGERS:

Up to two batteries (E2359, E2361, E2363, E2365, E2371, K0733) at any one time are allowed if required for a power wheelchair.

A non-sealed battery (E2358, E2360, E2362, E2364, E2372) will be denied as not reasonable and necessary.

A single mode battery charger (E2366) is appropriate for charging a sealed lead acid battery. If a dual mode battery charger (E2367) is provided as a replacement, it will be denied as not reasonable and necessary.

The usual maximum frequency of replacement for a lithium-based battery (E2397) is one every 3 years. Only one battery is allowed at any one time.

POWER TILT AND/OR RECLINE SEATING SYSTEMS (E1002-E1012):

A power seating system – tilt only, recline only, or combination tilt and recline – with or without power elevating legrests will be covered if criteria 1, 2, and 3 are met and if criterion 4, 5, or 6 is met:

1. The beneficiary meets all the coverage criteria for a power wheelchair described in the Power Mobility Devices LCD; and
2. A specialty evaluation that was performed by a licensed/certified medical professional, such as a physical therapist (PT) or occupational therapist (OT) or physician who has specific training and experience in rehabilitation wheelchair evaluations of the beneficiary's seating and positioning needs. The PT, OT, or physician may have no financial relationship with the supplier; and
3. The wheelchair is provided by a supplier that employs a RESNA-certified Assistive Technology Professional (ATP) who specializes in wheelchairs and who has direct, in-person involvement in the wheelchair selection for the beneficiary.
4. The beneficiary is at high risk for development of a pressure ulcer and is unable to perform a functional weight shift; or
5. The beneficiary utilizes intermittent catheterization for bladder management and is unable to independently transfer from the wheelchair to bed; or
6. The power seating system is needed to manage increased tone or spasticity.

If these criteria are not met, the power seating component(s) will be denied as not reasonable and necessary.

POWER WHEELCHAIR DRIVE CONTROL SYSTEMS:

An attendant control is covered in place of a beneficiary-operated drive control system if the beneficiary meets coverage criteria for a wheelchair, is unable to operate a manual or power wheelchair and has a caregiver who is unable to operate a manual wheelchair but is able to operate a power wheelchair.

OTHER POWER WHEELCHAIR ACCESSORIES:

An electronic interface (E2351) to allow a speech generating device to be operated by the power wheelchair control interface is covered if the beneficiary has a covered speech generating device. (Refer to the Speech Generating Devices LCD for details.)

MISCELLANEOUS ACCESSORIES:

Anti-rollback device (E0974) is covered if the beneficiary self-propels and needs the device because of ramps.

A safety belt/pelvic strap (E0978) is covered if the beneficiary has weak upper body muscles, upper body instability or muscle spasticity which requires use of this item for proper positioning.

One example (not all-inclusive) of a covered indication for swingaway, retractable, or removable hardware (E1028) would be to move the component out of the way so that a beneficiary can perform a slide transfer to a chair or bed.

A manual fully reclining back option (E1226) is covered if the beneficiary has one or more of the following conditions:

1. The beneficiary is at high risk for development of a pressure ulcer and is unable to perform a functional weight shift; or
2. The beneficiary utilizes intermittent catheterization for bladder management and is unable to independently transfer from the wheelchair to the bed.

If these criteria are not met, the manual reclining back will be denied as not reasonable and necessary.

For information concerning a push-rim activated power assist device for a manual wheelchair, refer to the Power Mobility Devices medical policy.

GENERAL

A detailed written order (DWO) (if applicable) must be received by the supplier before a claim is submitted. If the supplier bills for an item addressed in this policy without first receiving a completed DWO, the claim shall be denied as not reasonable and necessary.

For durable medical equipment, prosthetics, orthotics and supplies (DMEPOS) base items that require a Written Order Prior to Delivery (WOPD), the supplier must also obtain a DWO before submitting a claim for any associated options, accessories, and/or supplies that are separately billed. In this scenario, if the supplier bills for associated options, accessories, and/or supplies without first receiving a completed DWO, the claim shall be denied as not reasonable and necessary.

A WOPD (if applicable) must be received by the supplier before a DMEPOS item is delivered to a beneficiary. If a supplier delivers a DMEPOS item without first receiving a completed WOPD, the claim shall be statutorily denied. Refer to the LCD-related Policy Article, located at the bottom of this policy under the Related Local Coverage Documents section.

An item/service is correctly coded when it meets all the coding guidelines listed in CMS HCPCS guidelines, LCDs, LCD-related Policy Articles, or DME MAC articles. Claims that do not meet coding guidelines shall be denied as not reasonable and necessary/incorrectly coded.

Proof of delivery (POD) is a Supplier Standard and DMEPOS suppliers are required to maintain POD documentation in their files. Proof of delivery documentation must be made available to the Medicare contractor upon request. All services that do not have appropriate proof of delivery from the supplier shall be denied as not reasonable and necessary.

[Back to Top](#)

Coding Information

Bill Type Codes:

Contractors may specify Bill Types to help providers identify those Bill Types typically used to report this service. Absence of a Bill Type does not guarantee that the policy does not apply to that Bill Type. Complete absence of all Bill Types indicates that coverage is not influenced by Bill Type and the policy should be assumed to apply equally to all claims.

N/A

Revenue Codes:

Contractors may specify Revenue Codes to help providers identify those Revenue Codes typically used to report this service. In most instances Revenue Codes are purely advisory. Unless specified in the policy, services reported under other Revenue Codes are equally subject to this coverage determination. Complete absence of all Revenue Codes indicates that coverage is not influenced by Revenue Code and the policy should be assumed to apply equally to all Revenue Codes.

N/A

CPT/HCPCS Codes

Group 1 Paragraph:

The appearance of a code in this section does not necessarily indicate coverage.

HCPCS MODIFIERS:

EY – No physician or other licensed health care provider order for this item or service

GA - Waiver of liability statement issued, as required by payer policy, individual case

GY – Item or service statutorily excluded or does not meet the definition of any Medicare benefit

GZ - Item or service expected to be denied as not reasonable and necessary

KC - Replacement of special power wheelchair interface

KX - Requirements specified in the medical policy have been met

RB – Replacement of a part of DME furnished as part of a repair

HCPCS CODES:

ARM OF CHAIR:

Group 1 Codes:

E0973 WHEELCHAIR ACCESSORY, ADJUSTABLE HEIGHT, DETACHABLE ARMREST, COMPLETE ASSEMBLY, EACH
E2209 ACCESSORY, ARM TROUGH, WITH OR WITHOUT HAND SUPPORT, EACH
K0015 DETACHABLE, NON-ADJUSTABLE HEIGHT ARMREST, REPLACEMENT ONLY, EACH
K0017 DETACHABLE, ADJUSTABLE HEIGHT ARMREST, BASE, REPLACEMENT ONLY, EACH
K0018 DETACHABLE, ADJUSTABLE HEIGHT ARMREST, UPPER PORTION, REPLACEMENT ONLY, EACH
K0019 ARM PAD, REPLACEMENT ONLY, EACH
K0020 FIXED, ADJUSTABLE HEIGHT ARMREST, PAIR

Group 2 Paragraph: FOOTREST/LEGREST:

Group 2 Codes:

E0951 HEEL LOOP/HOLDER, ANY TYPE, WITH OR WITHOUT ANKLE STRAP, EACH
E0952 TOE LOOP/HOLDER, ANY TYPE, EACH
E0990 WHEELCHAIR ACCESSORY, ELEVATING LEG REST, COMPLETE ASSEMBLY, EACH
E0995 WHEELCHAIR ACCESSORY, CALF REST/PAD, REPLACEMENT ONLY, EACH
E1020 RESIDUAL LIMB SUPPORT SYSTEM FOR WHEELCHAIR, ANY TYPE
K0037 HIGH MOUNT FLIP-UP FOOTREST, REPLACEMENT ONLY, EACH
K0038 LEG STRAP, EACH
K0039 LEG STRAP, H STYLE, EACH
K0040 ADJUSTABLE ANGLE FOOTPLATE, EACH
K0041 LARGE SIZE FOOTPLATE, EACH
K0042 STANDARD SIZE FOOTPLATE, REPLACEMENT ONLY, EACH
K0043 FOOTREST, LOWER EXTENSION TUBE, REPLACEMENT ONLY, EACH
K0044 FOOTREST, UPPER HANGER BRACKET, REPLACEMENT ONLY, EACH
K0045 FOOTREST, COMPLETE ASSEMBLY, REPLACEMENT ONLY, EACH
K0046 ELEVATING LEGREST, LOWER EXTENSION TUBE, REPLACEMENT ONLY, EACH
K0047 ELEVATING LEGREST, UPPER HANGER BRACKET, REPLACEMENT ONLY, EACH
K0050 RATCHET ASSEMBLY, REPLACEMENT ONLY
K0051 CAM RELEASE ASSEMBLY, FOOTREST OR LEGREST, REPLACEMENT ONLY, EACH
K0052 SWINGAWAY, DETACHABLE FOOTRESTS, REPLACEMENT ONLY, EACH
K0053 ELEVATING FOOTRESTS, ARTICULATING (TELESCOPING), EACH

K0195 ELEVATING LEG RESTS, PAIR (FOR USE WITH CAPPED RENTAL WHEELCHAIR BASE)

Group 3 Paragraph: NONSTANDARD SEAT FRAME DIMENSIONS:

Group 3 Codes:

- E1011 MODIFICATION TO PEDIATRIC SIZE WHEELCHAIR, WIDTH ADJUSTMENT PACKAGE (NOT TO BE DISPENSED WITH INITIAL CHAIR)
- E2201 MANUAL WHEELCHAIR ACCESSORY, NONSTANDARD SEAT FRAME, WIDTH GREATER THAN OR EQUAL TO 20 INCHES AND LESS THAN 24 INCHES
- E2202 MANUAL WHEELCHAIR ACCESSORY, NONSTANDARD SEAT FRAME WIDTH, 24-27 INCHES
- E2203 MANUAL WHEELCHAIR ACCESSORY, NONSTANDARD SEAT FRAME DEPTH, 20 TO LESS THAN 22 INCHES
- E2204 MANUAL WHEELCHAIR ACCESSORY, NONSTANDARD SEAT FRAME DEPTH, 22 TO 25 INCHES
- K0056 SEAT HEIGHT LESS THAN 17" OR EQUAL TO OR GREATER THAN 21" FOR A HIGH STRENGTH, LIGHTWEIGHT, OR ULTRALIGHTWEIGHT WHEELCHAIR

Group 4 Paragraph: REAR WHEELS FOR MANUAL WHEELCHAIRS:

Group 4 Codes:

- E0961 MANUAL WHEELCHAIR ACCESSORY, WHEEL LOCK BRAKE EXTENSION (HANDLE), EACH
- E0967 MANUAL WHEELCHAIR ACCESSORY, HAND RIM WITH PROJECTIONS, ANY TYPE, REPLACEMENT ONLY, EACH
- E0988 MANUAL WHEELCHAIR ACCESSORY, LEVER-ACTIVATED, WHEEL DRIVE, PAIR
- E2205 MANUAL WHEELCHAIR ACCESSORY, HANDRIM WITHOUT PROJECTIONS (INCLUDES ERGONOMIC OR CONTOURED), ANY TYPE, REPLACEMENT ONLY, EACH
- E2206 MANUAL WHEELCHAIR ACCESSORY, WHEEL LOCK ASSEMBLY, COMPLETE, REPLACEMENT ONLY, EACH
- E2211 MANUAL WHEELCHAIR ACCESSORY, PNEUMATIC PROPULSION TIRE, ANY SIZE, EACH
- E2212 MANUAL WHEELCHAIR ACCESSORY, TUBE FOR PNEUMATIC PROPULSION TIRE, ANY SIZE, EACH
- E2213 MANUAL WHEELCHAIR ACCESSORY, INSERT FOR PNEUMATIC PROPULSION TIRE (REMOVABLE), ANY TYPE, ANY SIZE, EACH
- E2214 MANUAL WHEELCHAIR ACCESSORY, PNEUMATIC CASTER TIRE, ANY SIZE, EACH
- E2215 MANUAL WHEELCHAIR ACCESSORY, TUBE FOR PNEUMATIC CASTER TIRE, ANY SIZE, EACH
- E2216 MANUAL WHEELCHAIR ACCESSORY, FOAM FILLED PROPULSION TIRE, ANY SIZE, EACH
- E2217 MANUAL WHEELCHAIR ACCESSORY, FOAM FILLED CASTER TIRE, ANY SIZE, EACH
- E2218 MANUAL WHEELCHAIR ACCESSORY, FOAM PROPULSION TIRE, ANY SIZE, EACH
- E2219 MANUAL WHEELCHAIR ACCESSORY, FOAM CASTER TIRE, ANY SIZE, EACH
- E2220 MANUAL WHEELCHAIR ACCESSORY, SOLID (RUBBER/PLASTIC) PROPULSION TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2221 MANUAL WHEELCHAIR ACCESSORY, SOLID (RUBBER/PLASTIC) CASTER TIRE (REMOVABLE), ANY SIZE, REPLACEMENT ONLY, EACH
- E2222 MANUAL WHEELCHAIR ACCESSORY, SOLID (RUBBER/PLASTIC) CASTER TIRE WITH INTEGRATED WHEEL, ANY SIZE, REPLACEMENT ONLY, EACH
- E2224 MANUAL WHEELCHAIR ACCESSORY, PROPULSION WHEEL EXCLUDES TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2225 MANUAL WHEELCHAIR ACCESSORY, CASTER WHEEL EXCLUDES TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2226 MANUAL WHEELCHAIR ACCESSORY, CASTER FORK, ANY SIZE, REPLACEMENT ONLY, EACH
- E2227 MANUAL WHEELCHAIR ACCESSORY, GEAR REDUCTION DRIVE WHEEL, EACH
- E2228 MANUAL WHEELCHAIR ACCESSORY, WHEEL BRAKING SYSTEM AND LOCK, COMPLETE, EACH
- K0065 SPOKE PROTECTORS, EACH
- K0069 REAR WHEEL ASSEMBLY, COMPLETE, WITH SOLID TIRE, SPOKES OR MOLDED, REPLACEMENT ONLY, EACH
- K0070 REAR WHEEL ASSEMBLY, COMPLETE, WITH PNEUMATIC TIRE, SPOKES OR MOLDED, REPLACEMENT ONLY, EACH
- K0071 FRONT CASTER ASSEMBLY, COMPLETE, WITH PNEUMATIC TIRE, REPLACEMENT ONLY, EACH
- K0072 FRONT CASTER ASSEMBLY, COMPLETE, WITH SEMI-PNEUMATIC TIRE, REPLACEMENT ONLY, EACH
- K0073 CASTER PIN LOCK, EACH
- K0077 FRONT CASTER ASSEMBLY, COMPLETE, WITH SOLID TIRE, REPLACEMENT ONLY, EACH

Group 5 Paragraph: BATTERIES/CHARGERS:

Group 5 Codes:

- E2358 POWER WHEELCHAIR ACCESSORY, GROUP 34 NON-SEALED LEAD ACID BATTERY, EACH
- E2359 POWER WHEELCHAIR ACCESSORY, GROUP 34 SEALED LEAD ACID BATTERY, EACH (E.G., GEL CELL, ABSORBED GLASSMAT)
- E2360 POWER WHEELCHAIR ACCESSORY, 22NF NON-SEALED LEAD ACID BATTERY, EACH
- E2361 POWER WHEELCHAIR ACCESSORY, 22NF SEALED LEAD ACID BATTERY, EACH, (E.G., GEL CELL, ABSORBED GLASSMAT)
- E2362 POWER WHEELCHAIR ACCESSORY, GROUP 24 NON-SEALED LEAD ACID BATTERY, EACH
- E2363 POWER WHEELCHAIR ACCESSORY, GROUP 24 SEALED LEAD ACID BATTERY, EACH (E.G., GEL CELL, ABSORBED GLASSMAT)
- E2364 POWER WHEELCHAIR ACCESSORY, U-1 NON-SEALED LEAD ACID BATTERY, EACH
- E2365 POWER WHEELCHAIR ACCESSORY, U-1 SEALED LEAD ACID BATTERY, EACH (E.G., GEL CELL, ABSORBED GLASSMAT)
- E2366 POWER WHEELCHAIR ACCESSORY, BATTERY CHARGER, SINGLE MODE, FOR USE WITH ONLY ONE BATTERY TYPE, SEALED OR NON-SEALED, EACH
- E2367 POWER WHEELCHAIR ACCESSORY, BATTERY CHARGER, DUAL MODE, FOR USE WITH EITHER BATTERY TYPE, SEALED OR NON-SEALED, EACH
- E2371 POWER WHEELCHAIR ACCESSORY, GROUP 27 SEALED LEAD ACID BATTERY, (E.G., GEL CELL, ABSORBED GLASSMAT), EACH
- E2372 POWER WHEELCHAIR ACCESSORY, GROUP 27 NON-SEALED LEAD ACID BATTERY, EACH
- E2397 POWER WHEELCHAIR ACCESSORY, LITHIUM-BASED BATTERY, EACH
- K0733 POWER WHEELCHAIR ACCESSORY, 12 TO 24 AMP HOUR SEALED LEAD ACID BATTERY, EACH (E.G., GEL CELL, ABSORBED GLASSMAT)

Group 6 Paragraph: POWER SEATING SYSTEMS:**Group 6 Codes:**

- E1002 WHEELCHAIR ACCESSORY, POWER SEATING SYSTEM, TILT ONLY
- E1003 WHEELCHAIR ACCESSORY, POWER SEATING SYSTEM, RECLINE ONLY, WITHOUT SHEAR REDUCTION
- E1004 WHEELCHAIR ACCESSORY, POWER SEATING SYSTEM, RECLINE ONLY, WITH MECHANICAL SHEAR REDUCTION
- E1005 WHEELCHAIR ACCESSORY, POWER SEATING SYSTEM, RECLINE ONLY, WITH POWER SHEAR REDUCTION
- E1006 WHEELCHAIR ACCESSORY, POWER SEATING SYSTEM, COMBINATION TILT AND RECLINE, WITHOUT SHEAR REDUCTION
- E1007 WHEELCHAIR ACCESSORY, POWER SEATING SYSTEM, COMBINATION TILT AND RECLINE, WITH MECHANICAL SHEAR REDUCTION
- E1008 WHEELCHAIR ACCESSORY, POWER SEATING SYSTEM, COMBINATION TILT AND RECLINE, WITH POWER SHEAR REDUCTION
- E1009 WHEELCHAIR ACCESSORY, ADDITION TO POWER SEATING SYSTEM, MECHANICALLY LINKED LEG ELEVATION SYSTEM, INCLUDING PUSHROD AND LEG REST, EACH
- E1010 WHEELCHAIR ACCESSORY, ADDITION TO POWER SEATING SYSTEM, POWER LEG ELEVATION SYSTEM, INCLUDING LEG REST, PAIR
- E1012 WHEELCHAIR ACCESSORY, ADDITION TO POWER SEATING SYSTEM, CENTER MOUNT POWER ELEVATING LEG REST/PLATFORM, COMPLETE SYSTEM, ANY TYPE, EACH
- E2300 WHEELCHAIR ACCESSORY, POWER SEAT ELEVATION SYSTEM, ANY TYPE
- E2301 WHEELCHAIR ACCESSORY, POWER STANDING SYSTEM, ANY TYPE
- E2310 POWER WHEELCHAIR ACCESSORY, ELECTRONIC CONNECTION BETWEEN WHEELCHAIR CONTROLLER AND ONE POWER SEATING SYSTEM MOTOR, INCLUDING ALL RELATED ELECTRONICS, INDICATOR FEATURE, MECHANICAL FUNCTION SELECTION SWITCH, AND FIXED MOUNTING HARDWARE
- E2311 POWER WHEELCHAIR ACCESSORY, ELECTRONIC CONNECTION BETWEEN WHEELCHAIR CONTROLLER AND TWO OR MORE POWER SEATING SYSTEM MOTORS, INCLUDING ALL RELATED ELECTRONICS, INDICATOR FEATURE, MECHANICAL FUNCTION SELECTION SWITCH, AND FIXED MOUNTING HARDWARE

Group 7 Paragraph: POWER WHEELCHAIR DRIVE CONTROL SYSTEMS:**Group 7 Codes:**

- E2312 POWER WHEELCHAIR ACCESSORY, HAND OR CHIN CONTROL INTERFACE, MINI-PROPORTIONAL REMOTE JOYSTICK, PROPORTIONAL, INCLUDING FIXED MOUNTING HARDWARE
- E2313 POWER WHEELCHAIR ACCESSORY, HARNESS FOR UPGRADE TO EXPANDABLE CONTROLLER, INCLUDING ALL FASTENERS, CONNECTORS AND MOUNTING HARDWARE, EACH
- E2321

- POWER WHEELCHAIR ACCESSORY, HAND CONTROL INTERFACE, REMOTE JOYSTICK, NONPROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS, MECHANICAL STOP SWITCH, AND FIXED MOUNTING HARDWARE
- E2322 POWER WHEELCHAIR ACCESSORY, HAND CONTROL INTERFACE, MULTIPLE MECHANICAL SWITCHES, NONPROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS, MECHANICAL STOP SWITCH, AND FIXED MOUNTING HARDWARE
- E2323 POWER WHEELCHAIR ACCESSORY, SPECIALTY JOYSTICK HANDLE FOR HAND CONTROL INTERFACE, PREFABRICATED
- E2324 POWER WHEELCHAIR ACCESSORY, CHIN CUP FOR CHIN CONTROL INTERFACE
- E2325 POWER WHEELCHAIR ACCESSORY, SIP AND PUFF INTERFACE, NONPROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS, MECHANICAL STOP SWITCH, AND MANUAL SWINGAWAY MOUNTING HARDWARE
- E2326 POWER WHEELCHAIR ACCESSORY, BREATH TUBE KIT FOR SIP AND PUFF INTERFACE
- E2327 POWER WHEELCHAIR ACCESSORY, HEAD CONTROL INTERFACE, MECHANICAL, PROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS, MECHANICAL DIRECTION CHANGE SWITCH, AND FIXED MOUNTING HARDWARE
- E2328 POWER WHEELCHAIR ACCESSORY, HEAD CONTROL OR EXTREMITY CONTROL INTERFACE, ELECTRONIC, PROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS AND FIXED MOUNTING HARDWARE
- E2329 POWER WHEELCHAIR ACCESSORY, HEAD CONTROL INTERFACE, CONTACT SWITCH MECHANISM, NONPROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS, MECHANICAL STOP SWITCH, MECHANICAL DIRECTION CHANGE SWITCH, HEAD ARRAY, AND FIXED MOUNTING HARDWARE
- E2330 POWER WHEELCHAIR ACCESSORY, HEAD CONTROL INTERFACE, PROXIMITY SWITCH MECHANISM, NONPROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS, MECHANICAL STOP SWITCH, MECHANICAL DIRECTION CHANGE SWITCH, HEAD ARRAY, AND FIXED MOUNTING HARDWARE
- E2331 POWER WHEELCHAIR ACCESSORY, ATTENDANT CONTROL, PROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS AND FIXED MOUNTING HARDWARE
- E2373 POWER WHEELCHAIR ACCESSORY, HAND OR CHIN CONTROL INTERFACE, COMPACT REMOTE JOYSTICK, PROPORTIONAL, INCLUDING FIXED MOUNTING HARDWARE
- E2374 POWER WHEELCHAIR ACCESSORY, HAND OR CHIN CONTROL INTERFACE, STANDARD REMOTE JOYSTICK (NOT INCLUDING CONTROLLER), PROPORTIONAL, INCLUDING ALL RELATED ELECTRONICS AND FIXED MOUNTING HARDWARE, REPLACEMENT ONLY
- E2375 POWER WHEELCHAIR ACCESSORY, NON-EXPANDABLE CONTROLLER, INCLUDING ALL RELATED ELECTRONICS AND MOUNTING HARDWARE, REPLACEMENT ONLY
- E2376 POWER WHEELCHAIR ACCESSORY, EXPANDABLE CONTROLLER, INCLUDING ALL RELATED ELECTRONICS AND MOUNTING HARDWARE, REPLACEMENT ONLY
- E2377 POWER WHEELCHAIR ACCESSORY, EXPANDABLE CONTROLLER, INCLUDING ALL RELATED ELECTRONICS AND MOUNTING HARDWARE, UPGRADE PROVIDED AT INITIAL ISSUE

Group 8 Paragraph: OTHER POWER WHEELCHAIR ACCESSORIES:

Group 8 Codes:

- E1016 SHOCK ABSORBER FOR POWER WHEELCHAIR, EACH
- E1018 HEAVY DUTY SHOCK ABSORBER FOR HEAVY DUTY OR EXTRA HEAVY DUTY POWER WHEELCHAIR, EACH
- E2351 POWER WHEELCHAIR ACCESSORY, ELECTRONIC INTERFACE TO OPERATE SPEECH GENERATING DEVICE USING POWER WHEELCHAIR CONTROL INTERFACE
- E2368 POWER WHEELCHAIR COMPONENT, DRIVE WHEEL MOTOR, REPLACEMENT ONLY
- E2369 POWER WHEELCHAIR COMPONENT, DRIVE WHEEL GEAR BOX, REPLACEMENT ONLY
- E2370 POWER WHEELCHAIR COMPONENT, INTEGRATED DRIVE WHEEL MOTOR AND GEAR BOX COMBINATION, REPLACEMENT ONLY
- E2378 POWER WHEELCHAIR COMPONENT, ACTUATOR, REPLACEMENT ONLY
- E2381 POWER WHEELCHAIR ACCESSORY, PNEUMATIC DRIVE WHEEL TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2382 POWER WHEELCHAIR ACCESSORY, TUBE FOR PNEUMATIC DRIVE WHEEL TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2383 POWER WHEELCHAIR ACCESSORY, INSERT FOR PNEUMATIC DRIVE WHEEL TIRE (REMOVABLE), ANY TYPE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2384 POWER WHEELCHAIR ACCESSORY, PNEUMATIC CASTER TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2385 POWER WHEELCHAIR ACCESSORY, TUBE FOR PNEUMATIC CASTER TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2386 POWER WHEELCHAIR ACCESSORY, FOAM FILLED DRIVE WHEEL TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2387 POWER WHEELCHAIR ACCESSORY, FOAM FILLED CASTER TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
- E2388 POWER WHEELCHAIR ACCESSORY, FOAM DRIVE WHEEL TIRE, ANY SIZE, REPLACEMENT ONLY, EACH

E2389 POWER WHEELCHAIR ACCESSORY, FOAM CASTER TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
E2390 POWER WHEELCHAIR ACCESSORY, SOLID (RUBBER/PLASTIC) DRIVE WHEEL TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
E2391 POWER WHEELCHAIR ACCESSORY, SOLID (RUBBER/PLASTIC) CASTER TIRE (REMOVABLE), ANY SIZE, REPLACEMENT ONLY, EACH
E2392 POWER WHEELCHAIR ACCESSORY, SOLID (RUBBER/PLASTIC) CASTER TIRE WITH INTEGRATED WHEEL, ANY SIZE, REPLACEMENT ONLY, EACH
E2394 POWER WHEELCHAIR ACCESSORY, DRIVE WHEEL EXCLUDES TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
E2395 POWER WHEELCHAIR ACCESSORY, CASTER WHEEL EXCLUDES TIRE, ANY SIZE, REPLACEMENT ONLY, EACH
E2396 POWER WHEELCHAIR ACCESSORY, CASTER FORK, ANY SIZE, REPLACEMENT ONLY, EACH
K0098 DRIVE BELT FOR POWER WHEELCHAIR, REPLACEMENT ONLY

Group 9 Paragraph: MISCELLANEOUS ACCESSORIES:

Group 9 Codes:

A9270 NON-COVERED ITEM OR SERVICE
A9900 MISCELLANEOUS DME SUPPLY, ACCESSORY, AND/OR SERVICE COMPONENT OF ANOTHER HCPCS CODE
E0705 TRANSFER DEVICE, ANY TYPE, EACH
E0950 WHEELCHAIR ACCESSORY, TRAY, EACH
E0958 MANUAL WHEELCHAIR ACCESSORY, ONE-ARM DRIVE ATTACHMENT, EACH
E0959 MANUAL WHEELCHAIR ACCESSORY, ADAPTER FOR AMPUTEE, EACH
E0971 MANUAL WHEELCHAIR ACCESSORY, ANTI-TIPPING DEVICE, EACH
E0974 MANUAL WHEELCHAIR ACCESSORY, ANTI-ROLLBACK DEVICE, EACH
E0978 WHEELCHAIR ACCESSORY, POSITIONING BELT/SAFETY BELT/PELVIC STRAP, EACH
E0981 WHEELCHAIR ACCESSORY, SEAT UPHOLSTERY, REPLACEMENT ONLY, EACH
E0982 WHEELCHAIR ACCESSORY, BACK UPHOLSTERY, REPLACEMENT ONLY, EACH
E0985 WHEELCHAIR ACCESSORY, SEAT LIFT MECHANISM
E1014 RECLINING BACK, ADDITION TO PEDIATRIC SIZE WHEELCHAIR
E1015 SHOCK ABSORBER FOR MANUAL WHEELCHAIR, EACH
E1017 HEAVY DUTY SHOCK ABSORBER FOR HEAVY DUTY OR EXTRA HEAVY DUTY MANUAL WHEELCHAIR, EACH
E1028 WHEELCHAIR ACCESSORY, MANUAL SWINGAWAY, RETRACTABLE OR REMOVABLE MOUNTING HARDWARE FOR JOYSTICK, OTHER CONTROL INTERFACE OR POSITIONING ACCESSORY
E1029 WHEELCHAIR ACCESSORY, VENTILATOR TRAY, FIXED
E1030 WHEELCHAIR ACCESSORY, VENTILATOR TRAY, GIMBALED
E1225 WHEELCHAIR ACCESSORY, MANUAL SEMI-RECLINING BACK, (RECLINE GREATER THAN 15 DEGREES, BUT LESS THAN 80 DEGREES), EACH
E1226 WHEELCHAIR ACCESSORY, MANUAL FULLY RECLINING BACK, (RECLINE GREATER THAN 80 DEGREES), EACH
E2207 WHEELCHAIR ACCESSORY, CRUTCH AND CANE HOLDER, EACH
E2208 WHEELCHAIR ACCESSORY, CYLINDER TANK CARRIER, EACH
E2210 WHEELCHAIR ACCESSORY, BEARINGS, ANY TYPE, REPLACEMENT ONLY, EACH
E2230 MANUAL WHEELCHAIR ACCESSORY, MANUAL STANDING SYSTEM
E2295 MANUAL WHEELCHAIR ACCESSORY, FOR PEDIATRIC SIZE WHEELCHAIR, DYNAMIC SEATING FRAME, ALLOWS COORDINATED MOVEMENT OF MULTIPLE POSITIONING FEATURES
K0105 IV HANGER, EACH
K0108 WHEELCHAIR COMPONENT OR ACCESSORY, NOT OTHERWISE SPECIFIED

ICD-10 Codes that Support Medical Necessity

Group 1 Paragraph: Not specified

Group 1 Codes: N/A

ICD-10 Codes that DO NOT Support Medical Necessity

Group 1 Paragraph: Not specified

Group 1 Codes: N/A

ICD-10 Additional Information

[Back to Top](#)

General Information

Associated Information

DOCUMENTATION REQUIREMENTS

Section 1833(e) of the Social Security Act precludes payment to any provider of services unless "there has been furnished such information as may be necessary in order to determine the amounts due such provider." It is expected that the beneficiary's medical records will reflect the need for the care provided. The beneficiary's medical records include the physician's office records, hospital records, nursing home records, home health agency records, records from other healthcare professionals and test reports. This documentation must be available upon request.

GENERAL DOCUMENTATION REQUIREMENTS

In order to justify payment for DMEPOS items, suppliers must meet the following requirements:

- Prescription (orders)
- Medical Record Information (including continued need/use if applicable)
- Correct Coding
- Proof of Delivery

Refer to the LCD-related Standard Documentation Requirements article, located at the bottom of this policy under the Related Local Coverage Documents section for additional information regarding these requirements.

Refer to the Supplier Manual for additional information on documentation requirements.

Refer to the DME MAC web sites for additional bulletin articles and other publications related to this LCD.

POLICY SPECIFIC DOCUMENTATION REQUIREMENTS

Items covered in this LCD have additional policy-specific requirements that must be met prior to Medicare reimbursement.

Refer to the LCD-related Policy article, located at the bottom of this policy under the Related Local Coverage Documents section for additional information.

Miscellaneous

Appendices

Utilization Guidelines

Refer to Coverage Indications, Limitations and/or Medical Necessity.

Sources of Information and Basis for Decision

Reserved for future use [Back to Top](#)

Revision History Information

Revision History Date	Revision History Number	Revision History Explanation	Reason(s) for Change
01/01/2017	R5	<p>Revision Effective Date: 01/01/2017 COVERAGE INDICATIONS, LIMITATIONS AND/OR MEDICAL NECESSITY: Removed: Standard Documentation Language Added: New reference language and Directions to Standard Documentation Requirements Added: General Requirements HCPCS CODES: Revised: Descriptor changed in Group 1 - K0015, K0019; Group 2 - E0995, K0037, K0042, K0043, K0044, K0045, K0046, K0047, K0050, K0051, K0052; Group 4 -E0967, E2206, E2220, E2221, E2222, E2224, K0069, K0070, K0071, K0072, K0077; Group 8 - K0098 DOCUMENTATION REQUIREMENTS: Removed: Standard Documentation Language Added: General Documentation Requirements Added: New reference language and Directions to Standard Documentation Requirements POLICY SPECIFIC DOCUMENTATION REQUIREMENTS: Removed: Standard Documentation Language Added: Direction to Standard Documentation Requirements Removed: Information under Miscellaneous Removed: Information under Appendices RELATED LOCAL COVERAGE DOCUMENTS: Added: LCD-related Standard Documentation Requirements article</p>	<ul style="list-style-type: none"> • Provider Education/Guidance • Revisions Due To ICD-10-CM Code Changes
07/01/2016	R4	<p>Effective July 1, 2016 oversight for DME MAC LCDs is the responsibility of CGS Administrators, LLC 18003 and 17013 and Noridian Healthcare Solutions, LLC 19003 and 16013. No other changes have been made to the LCDs.</p>	<ul style="list-style-type: none"> • Change in Assigned States or Affiliated Contract Numbers
01/01/2016	R3	<p>Revision Effective Date: 01/01/2016 COVERAGE INDICATIONS, LIMITATIONS AND/OR MEDICAL NECESSITY: Added: HCPCS code E1012 to Power Tilt and/or Recline Seating Systems range HCPCS CODES: Added: HCPCS code E1012 Revised: K0017 and K0018 long narrative description DOCUMENTATION REQUIREMENTS: Revised: Standard Documentation Language to remove start date verbiage from Prescription Requirements (Effective 11/5/2015)</p>	<ul style="list-style-type: none"> • Provider Education/Guidance • Revisions Due To CPT/HCPCS Code Changes
10/01/2015	R2	<p>Revision Effective Date: 10/31/2014 COVERAGE INDICATIONS, LIMITATIONS AND/OR MEDICAL NECESSITY: Revised: Standard Documentation Language to add covered prior to a beneficiary's Medicare eligibility Added: HCPCS Codes E2358 and E2359 to the Batteries/Chargers section DOCUMENTATION REQUIREMENTS: Deleted: Reference to refill of supplies from Continued Use Revised: Standard Documentation Language to add who can enter date of delivery date on the POD Added: Repair/Replacement section 07/01/2013 (June 2014 Publication)</p>	<ul style="list-style-type: none"> • Provider Education/Guidance
10/01/2015	R1	<p>MISCELLANEOUS: Removed: Requirement for accessories to be billed on the same claim as base</p>	<ul style="list-style-type: none"> • Provider Education/Guidance

[Back to Top](#)

[Associated Documents](#)

Attachments N/A

Related Local Coverage Documents Article(s) [A55426 - Standard Documentation Requirements for All Claims Submitted to DME MACs A52504 - Wheelchair Options/Accessories - Policy Article](#)

Related National Coverage Documents N/A

Public Version(s) Updated on 03/09/2017 with effective dates 01/01/2017 - N/A [Updated on 06/07/2016 with effective dates 07/01/2016 - 12/31/2016](#) [Updated on 02/26/2016 with effective dates 01/01/2016 - 06/30/2016](#)
Some older versions have been archived. Please visit the [MCD Archive Site](#) to retrieve them. [Back to Top](#)

Keywords

N/A Read the [LCD Disclaimer](#) [Back to Top](#)

END OF LOCAL COVERAGE DETERMINATION

Per the Code of Federal Regulations, 42 C.F.R § 426. 325, only those portions of the currently effective Local Coverage Determination (LCD) that are based on section 1862(a)(1)(A) of the Social Security Act, may be challenged through an acceptable complaint as described in 42 C.F.R § 426.400. Also, per 42 C.F.R § 426.325 items that are not reviewable, and therefore cannot be challenged, include the Policy Article. Please note the distinction of the documents when reviewing the materials.

Local Coverage Article: Wheelchair Options/Accessories - Policy Article (A52504)

Links in PDF documents are not guaranteed to work. To follow a web link, please use the MCD Website.

Contractor Information

Contractor Name	Contract Type	Contract Number	Jurisdiction	State(s)
CGS Administrators, LLC	DME MAC	17013 -	DME MAC J-B	Illinois Indiana Kentucky Michigan Minnesota Ohio Wisconsin Alabama Arkansas Colorado Florida Georgia Louisiana Mississippi North Carolina New Mexico Oklahoma Puerto Rico South Carolina Tennessee Texas Virginia Virgin Islands West Virginia Connecticut District of Columbia Delaware Massachusetts Maryland Maine New Hampshire New Jersey New York - Entire State Pennsylvania Rhode Island Vermont Alaska American Samoa Arizona California - Entire State Guam Hawaii Iowa Idaho Kansas Missouri - Entire State Montana North Dakota Nebraska Nevada Oregon South Dakota
CGS Administrators, LLC	DME MAC	18003 -	DME MAC J-C	
Noridian Healthcare Solutions, LLC	DME MAC	16013 -	DME MAC J-A	
Noridian Healthcare Solutions, LLC	DME MAC	19003 -	DME MAC J-D	

Article Information

General Information

Article ID

A52504

Original ICD-9 Article ID

[A47229](#)[A20284](#)[A19829](#)[A19846](#)**Original Article Effective Date**

10/01/2015

Revision Effective Date

01/01/2017

Revision Ending Date

N/A

Retirement Date

N/A

Article Title

Wheelchair Options/Accessories - Policy Article

AMA CPT / ADA CDT / AHA NUBC Copyright Statement

CPT only copyright 2002-2017 American Medical Association. All Rights Reserved. CPT is a registered trademark of the American Medical Association. Applicable FARS/DFARS Apply to Government Use. Fee schedules, relative value units, conversion factors and/or related components are not assigned by the AMA, are not part of CPT, and the AMA is not recommending their use. The AMA does not directly or indirectly practice medicine or dispense medical services. The AMA assumes no liability for data contained or not contained herein.

The Code on Dental Procedures and Nomenclature (Code) is published in Current Dental Terminology (CDT). Copyright © American Dental Association. All rights reserved. CDT and CDT-2016 are trademarks of the American Dental Association.

UB-04 Manual. OFFICIAL UB-04 DATA SPECIFICATIONS MANUAL, 2014, is copyrighted by American Hospital Association ("AHA"), Chicago, Illinois. No portion of OFFICIAL UB-04 MANUAL may be reproduced, sorted in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, recording or otherwise, without prior express, written consent of AHA." Health Forum reserves the right to change the copyright notice from time to time upon written notice to Company.

Article Guidance

Article Text:**NON-MEDICAL NECESSITY COVERAGE AND PAYMENT RULES:**

For any item to be covered by Medicare, it must 1) be eligible for a defined Medicare benefit category, 2) be reasonable and necessary for the diagnosis or treatment of illness or injury or to improve the functioning of a malformed body member, and 3) meet all other applicable Medicare statutory and regulatory requirements. Information provided in this policy article relates to determinations other than those based on Social Security Act §1862(a)(1)(A) provisions (i.e. "reasonable and necessary").

Wheelchair options and accessories are covered under the Durable Medical Equipment benefit (Social Security Act §1861(s)(6)). In order for a beneficiary's equipment to be eligible for reimbursement the reasonable and necessary (R&N) requirements set out in the related Local Coverage Determination must be met. In addition, there are specific statutory payment policy requirements, discussed below, that also must be met.

The allowance for a power operated vehicle (POV) includes all options and accessories that are provided at the time of initial issue, including but not limited to batteries, battery chargers, seating systems, etc. If a beneficiary-owned POV meets coverage criteria, medically necessary replacement items are covered.

The allowance for a rollabout chair includes all options and accessories that are provided at the time of initial issue. The allowance for a transport chair includes all options and accessories that are provided at the time of initial issue except for elevating legrests (E0990, K0195). If a rollabout chair or transport chair are covered, medically necessary replacement items are covered.

An option/accessory that is beneficial primarily in allowing the beneficiary to perform leisure or recreational activities is non-covered.

If an option or accessory that is included in another code is billed separately, the claim line will be denied as not separately payable. (Refer to Coding Guidelines section for additional information on correct coding.)

BATTERIES/ CHARGERS:

A sealed battery (E2359, E2361, E2363, E2365, E2371, E2397, K0733) is separately payable from a power wheelchair base.

There is no additional/separate payment when a dual mode battery charger is provided at the time of initial issue of a power wheelchair.

A battery charger (E2366, E2367) is included in the allowance for a power wheelchair base.

POWER SEATING SYSTEMS:

A power seat elevation feature (E2300) and power standing feature (E2301) are non-covered because they are not primarily medical in nature. If a wheelchair has an electrical connection device described by code E2310 or E2311 and if the sole function of the connection is for a power seat elevation or power standing feature, it will be denied as non-covered.

POWER WHEELCHAIR DRIVE CONTROL SYSTEMS:

If an attendant control (E2331) is provided in addition to a beneficiary-operated drive control system, it will be denied as non-covered. (See the related LCD for situations in which it is provided in place of a beneficiary-operated system.)

OTHER POWER WHEELCHAIR ACCESSORIES:

An electronic interface used to control lights or other electrical devices is non-covered because it is not primarily medical in nature.

The following features of a power wheelchair will be denied as non-covered: stair climbing (A9270), electronic balance (A9270), ability to elevate the seat by balancing on two wheels (A9270), and remote operation (A9270).

MISCELLANEOUS ACCESSORIES:

Swingaway, retractable, or removable hardware (E1028) is non-covered if the primary indication for its use is to allow the beneficiary to move close to desks or other surfaces. If it ordered for this indication, a GY modifier must be added to the code.

A manual standing system for a manual wheelchair (E2230) is non-covered (no benefit category) because it is not primarily medical in nature.

POLICY SPECIFIC DOCUMENTATION REQUIREMENTS

In addition to policy specific documentation requirements, there are general documentation requirements that are applicable to all DMEPOS policies. These general requirements are located in the DOCUMENTATION REQUIREMENTS section of the LCD.

Refer to the LCD-related Standard Documentation Requirements article, located at the bottom of this Policy Article under the Related Local Coverage Documents section for additional information regarding GENERAL DOCUMENTATION REQUIREMENTS and the POLICY SPECIFIC DOCUMENTATION REQUIREMENTS discussed below.

REQUIREMENTS FOR SPECIFIC DMEPOS ITEMS PURSUANT TO 42 CFR 410.38(g)

42 CFR 410.38(g) requires a face-to-face evaluation and a specific written order prior to delivery for specified HCPCS codes. CMS provides a list of the specified codes, which is periodically updated and located [here](#), on CMS's website.

Claims for the specified items subject to 42 CFR 410.38(g) that do not meet the requirements specified in the LCD-related Standard Documentation Requirements Article will be denied as statutorily noncovered – failed to meet statutory requirements.

If the supplier delivers the item prior to receipt of a written order, it will be denied as statutorily noncovered. If the written order is not obtained prior to delivery, payment will not be made for that item even if a written order is subsequently obtained. If a similar item is subsequently provided by an unrelated supplier who has obtained a written order prior to delivery, it will be eligible for coverage.

MODIFIERS

KX, GA, GY, AND GZ MODIFIERS:

For accessories for a power mobility device, if the requirements related to a 7-element order and face-to-face examination in the Power Mobility Devices Policy Article have not been met, the GY modifier must be added to the codes for all accessories.

For accessories provided with a manual wheelchair or power mobility device, if it is only needed for mobility outside the home, the GY modifier must be added to the codes for all accessories.

If the conditions for use of the GY modifier are not met, the KX modifier must be added to the code for the accessory only if (a) the coverage criteria that are specified in the Manual Wheelchair Bases or Power Mobility Devices LCD have been met and (b) any specific coverage criteria for the accessory in the related LCD have been met. If the coverage criteria are not met, the KX modifier must not be used.

If the conditions for use of the GY modifier are not met and if the requirements for use of the KX modifier are not met, the GA or GZ modifier must be added to a claim line for the accessory. When there is an expectation of a medical necessity denial, suppliers must enter the GA modifier on the claim line if they have obtained a properly executed Advance Beneficiary Notice (ABN) or the GZ modifier if they have not obtained a valid ABN.

If the GY modifier is used, the KX, GA, and GZ modifiers should not be used.

Claim lines billed without a GA, GY, GZ, or KX modifier will be rejected as missing information.

CODING GUIDELINES

GENERAL:

Power Wheelchair Basic Equipment Package - Each power wheelchair code is required to include all these items on initial issue (i.e., no separate billing/payment at the time of initial issue, unless otherwise noted). The statement that an item may be separately billed does not necessarily indicate coverage.

- Lap belt or safety belt. Shoulder harness/straps or chest straps/vest may be billed separately.
- Battery charger, single mode
- Complete set of tires and casters, any type
- Legrests. There is no separate billing/payment if fixed, swingaway, or detachable non-elevating legrests with or without calf pad are provided. Elevating legrests may be billed separately.

- Footrests/foot platform. There is no separate billing/payment if fixed, swingaway, or detachable footrests or a foot platform without angle adjustment are provided. There is no separate billing for angle adjustable footplates with Group 1 or 2 PWCs. Angle adjustable footplates may be billed separately with Group 3, 4 and 5 PWCs.
- Armrests. There is no separate billing/ payment if fixed, swingaway, or detachable non-adjustable height armrests with arm pad are provided (K0015). Adjustable height armrests (E0973, K0020) may be billed separately.
- Any weight specific components (braces, bars, upholstery, brackets, motors, gears, etc.) as required by beneficiary weight capacity.
- Any seat width and depth. Exception: For Group 3 and 4 PWCs with a sling/solid seat/back, the following may be billed separately:
 - For Standard Duty, seat width and/or depth greater than 20 inches;
 - For Heavy Duty, seat width and/or depth greater than 22 inches;
 - For Very Heavy Duty, seat width and/or depth greater than 24 inches;
 - For Extra Heavy Duty, no separate billing
- Any back width. Exception: For Group 3 and 4 PWCs with a sling/solid seat/back, the following may be billed separately:
 - For Standard Duty, back width greater than 20 inches;
 - For Heavy Duty, back width greater than 22 inches;
 - For Very Heavy Duty, back width greater than 24 inches;
 - For Extra Heavy Duty, no separate billing
- Controller and Input Device. There is no separate billing/payment if a non-expandable controller and a standard proportional joystick (integrated or remote) is provided. An expandable controller, a nonstandard joystick (i.e., non-proportional or mini, compact or short throw proportional), or other alternative control device may be billed separately.

Power Operated Vehicle (POV) Basic Equipment Package - Each POV is to include all these items on initial issue (i.e., no separate billing/payment at time of initial issue):

- Lap belt or safety belt. Shoulder harness/straps or chest straps/vest may be billed separately
- Battery or batteries required for operation
- Battery charger, single mode
- Weight appropriate upholstery and seating system
- Tiller steering
- Non-expandable controller with proportional response to input
- Complete set of tires
- All accessories needed for safe operation

A table at the end of this section defines the bundling guidelines for wheelchair bases and options/accessories. Codes listed in Column II are not separately payable from the wheelchair base and must not be billed separately at the time of initial purchase or rental of the wheelchair.

Codes K0015, K0019, K0037, K0042, K0043, K0044, K0045, K0046, K0047, K0050, K0051, K0052, K0069, K0070, K0071, K0072, K0077, K0098, E0967, E0995, E2206, E2220, E2221, E2222, and E2224 are for replacement items only. These codes are not used at the time of initial issue.

A replacement option/accessory for POV is billed using a wheelchair option/accessory code. All options and accessories provided at the time of initial issue of a POV are not separately billable.

Accessories provided at the time of initial issue of a rollabout chair are not separately billable. Accessories provided with the initial issue of a transport chair are not separately billable with the exception of elevating legrests (E0990, K0195). A replacement accessory for a rollabout or transport chair is billed using code E1399.

The RB modifier is used when an option or accessory is provided as a replacement for the same part which has been worn or damaged (e.g., replacing a tire of the same type). The RB modifier must not be used for an upgrade subsequent to providing the wheelchair base (e.g., replacing a standard seat of a power wheelchair with a power seating system). The RB modifier must not be used if the accessory is provided at the same time as the wheelchair base, even if the option/accessory is the same as one that the beneficiary had on a prior wheelchair. (See section on Power Wheelchair Drive Control Systems for instructions on the use of the KC replacement modifier.)

Miscellaneous options, accessories, or replacement parts for wheelchairs that do not have a specific HCPCS code and are not included in another code should be coded K0108. If multiple miscellaneous accessories are provided, each should be billed on a separate claim line using code K0108. When billing more than one line item with code K0108, ensure that the additional information can be matched to the appropriate line item on the claim. It is also helpful to reference the line item to the submitted charge. If a supplier chooses to bill separately for a component that is included in another code, code A9900 must be used.

The right (RT) and left (LT) modifiers must be used when appropriate. If bilateral items (left and right) are provided as a purchase and the unit of service of the code is "each", bill both items on the same claim line using the LTRT modifiers and 2 units of service. If bilateral items are provided as a rental and the unit of service is "each", bill the items on two separate claim lines with the RT modifier on one line and the LT modifier on the other. If bilateral items are provided and the unit of service is "pair", the LT and RT modifiers do not need to be reported.

Codes E0968, E0969, E0970, E0980, E0994, E1227, E1228, E1296-E1298, and E2340-E2343 are not valid for claim submission.

Suppliers should contact the Pricing, Data Analysis, and Coding (PDAC) contractor for guidance on correct coding of these items.

FOOTREST/ LEGREST:

Elevating legrests that are used with a wheelchair that is purchased or owned by the beneficiary are coded E0990. This code is per legrest. Elevating legrests that are used with a capped rental wheelchair base are coded K0195. This code is per pair of legrests.

NONSTANDARD SEAT FRAME DIMENSIONS:

For all adult manual wheelchairs (E1161, K0001-K0009), payment for seat widths and/or seat depths of 15-19 inches is included in the payment for the base code. These seat dimensions should not be billed separately. Codes E2201-E2204 describe seat widths and/or depths of 20 inches or more for manual wheelchairs.

For power wheelchairs, there is no separate billing for nonstandard seat frame dimensions (width, depth, or height) with the following exceptions: For Group 3 and 4 power wheelchairs, with a sling/solid seat/back, the following items may be billed separately using code K0108:

- For Standard Duty, seat width and/or depth greater than 20 inches;
- For Heavy Duty, seat width and/or depth greater than 22 inches;
- For Very Heavy Duty, seat width and/or depth greater than 24 inches;
- For Extra Heavy Duty, no separate billing

For Group 3 and 4 PWCs with a sling/solid seat/back, the following items may be billed separately using code K0108:

- For Standard Duty, back width greater than 20 inches;
- For Heavy Duty, back width greater than 22 inches;
- For Very Heavy Duty, back width greater than 24 inches;
- For Extra Heavy Duty, no separate billing

Code K0108 may not be billed for nonstandard dimensions of a power tilt and/or recline seating system (E1002-E1008). The definition of those codes includes any frame width and depth.

WHEELS/TIRES FOR MANUAL WHEELCHAIRS:

A propulsion wheel is a large wheel which can be used by a beneficiary to propel the wheelchair with his/her arms.

A caster is a small wheel that is in contact with the ground during normal operation of the wheelchair and which cannot be used for arm propulsion. This includes rear tires on tilt-in-space wheelchairs that are not used for arm propulsion.

A lever activated drive (E0988) is an alternative drive mechanism for propulsion of a manual wheelchair. It includes a user-powered lever-arm mechanism attached to one or both wheel hub(s). The lever activates adjustable-ratio gears and has the capability to shift between forward, reverse and braking.

A pneumatic tire (E2211, E2214) is a rubber tire which is used in conjunction with a separate tube (E2212, E2215) which is filled with air.

A flat free insert (E2213) is a removable ring of firm material that is placed inside of a pneumatic tire to allow the wheelchair to continue to move if the pneumatic tire is punctured. This code may not be used for a foam filled tire.

A foam filled tire (E2216, E2217) is one in which a rubber tire shell has been filled with foam which is non-removable.

A foam tire (E2218, E2219) is one which is made entirely of self-skinning urethane.

A replacement only solid tire (E2220, E2221, E2222) is one which is made of hard plastic or rubber.

A gear reduction drive wheel (E2227) is one that has more than one gear ratio option. Pushing on the rim allows the user to manually shift between the gears in order to provide additional leverage to assist propulsion of a manual wheelchair.

A wheel braking and lock system (E2228) is a caliper or disc type braking system that permits the controlled slowing of a manual wheelchair or the controlled descent on inclines. It also has full wheel lock capability.

A replacement only rear wheel assembly (K0069, K0070) includes a wheel rim plus a tire. For pneumatic tires, it also includes the tire tube, but not a flat free insert.

A replacement only caster assembly (K0071, K0072, K0077) includes a caster fork, wheel rim, and tire.

For information concerning a push-rim activated power assist device for a manual wheelchair, refer to the Power Mobility Devices medical policy.

POWER SEATING SYSTEMS:

A power tilt seating system (E1002) includes: a solid seat platform and a solid back; any frame width and depth; detachable or flip-up fixed height or adjustable height armrests; fixed or swingaway detachable legrests; fixed or flip-up footplates; a motor and related electronics with or without variable speed programmability; a switch control which is independent of the power wheelchair drive control interface; any hardware that is needed to attach the seating system to the wheelchair base. It does not include a headrest. It must have the following features: ability to tilt to greater than or equal to 20 degrees from horizontal; back height of at least 20 inches; ability for the supplier to adjust the seat to back angle; ability to support beneficiary weight of at least 250 pounds.

A power recline seating system (E1003-E1005) includes: a solid seat platform and a solid back; any frame width and depth; detachable or flip-up fixed height or adjustable height arm rests; fixed or swingaway detachable legrests; fixed or flip-up footplates; a motor and related electronics with or without variable speed programmability; a switch control which is independent of the power wheelchair drive control interface; any hardware that is needed to attach the seating system to the wheelchair base. It does not include a headrest. It must have the following features: ability to recline to greater than or equal to 150 degrees from horizontal; back height of at least 20 inches; ability to support beneficiary weight of at least 250 pounds.

A power tilt and recline seating system (E1006-E1008) includes: a solid seat platform and a solid back; any frame width and depth; detachable or flip-up fixed height or adjustable height armrests; fixed or swingaway detachable legrests; fixed or flip-up footplates; two motors and related electronics with or without variable speed programmability; a switch control which is independent of the power wheelchair drive control interface; any hardware that is needed to attach the seating system to the wheelchair base. It does not include a headrest. It must have the following features: ability to tilt to greater than or equal to 20 degrees from horizontal; ability to recline to greater than or equal to 150 degrees from horizontal; back height of at least 20 inches; ability to support beneficiary weight of at least 250 pounds.

Coding for a power tilt system (E1002), power recline system (E1003, E1004 and E1005), and tilt/recline system (E1006, E1007 and E1008) are all-inclusive. Usage of K0108 to bill for additional heavy duty or bariatric features is considered unbundling and is not allowed.

A power tilt seating system or power tilt and recline seating system which does not achieve a tilt of greater than or equal to 20 degrees is considered to be the same as the standard seat included in the base wheelchair. Codes E1002 – E1008 must not be used to describe a power tilt seating system or a power tilt and recline seating system which does not achieve a tilt of greater than or equal to 20 degrees. These seating systems must be coded as A9900 and are not separately payable.

A mechanical shear reduction feature (E1004 and E1007) consists of two separate back panels. As the posterior back panel reclines or raises there is a mechanical linkage between the two panels which allows the beneficiary's back to stay in contact with the anterior panel without sliding along that panel.

A power shear reduction feature (E1005 and E1008) consists of two separate back panels. As the posterior back panel reclines or raises there is a separate motor which controls the linkage between the two panels and allows the beneficiary's back to stay in contact with the anterior panel without sliding along that panel.

A mechanically linked leg elevation feature (E1009) involves a pushrod which connects the legrest to a power recline seating system. With this feature, when the back reclines, the legrest elevates; when the back raises, the legrest lowers.

A power leg elevation feature (E1010, E1012) involves dedicated motor(s) and related electronics with or without variable speed programmability which allows the legrest to be raised and lowered independently of the recline and/or tilt of the seating system. It includes a switch control which may or may not be integrated with the power tilt and/or recline control(s). It includes either articulating or non-articulating legrests. The unit of service of code E1010 is a pair. The unit of service for code E1012 is each.

A power seat elevation system (E2300) includes: a motor and related electronics with or without variable speed programmability; a switch control which is independent of the power wheelchair drive control interface; any hardware that is needed to attach the seating system to the wheelchair base. It must provide a seat elevation of at least 6 inches.

A power standing system (E2301) includes: a solid seat platform and a solid back; detachable or flip-up fixed height armrests; hinged legrests; anterior knee supports; fixed or flip-up footplates; a motor and related electronics with or without variable speed programmability; a basic switch control which is independent of the power wheelchair drive control interface; any hardware that is needed to attach the seating system to the wheelchair base. It does not include a headrest. It must have the following features: ability to move the beneficiary to a standing position; ability to support beneficiary weight of at least 250 pounds.

Codes E2310 and E2311 describe the electronic components that allow the beneficiary to control two or more of the following motors from a single interface (e.g., proportional joystick, touchpad, or non-proportional interface): power wheelchair drive, power tilt, power recline, power shear reduction, power leg elevation, power seat elevation, power standing. It includes a function selection switch which allows the beneficiary to select the motor that is being controlled and an indicator feature to visually show which function has been selected. When the wheelchair drive function has been selected, the indicator feature may also show the direction that has been selected (forward, reverse, left, right). This indicator feature may be in a separate display box or may be integrated into the wheelchair interface. Payment for the code includes an allowance for fixed mounting hardware for the control box and for the display box (if present).

POWER WHEELCHAIR DRIVE CONTROL SYSTEMS:

The term interface in the code narrative and definitions describes the mechanism for controlling the movement of a power wheelchair. Examples of interfaces include, but are not limited to, joystick, sip and puff, chin control, head control, etc. (Note: In the Power Mobility Devices policy, the term "control input device" is used instead of "interface".)

A proportional interface is one in which the direction and amount of movement by the beneficiary controls the direction and speed of the wheelchair. One example of a proportional interface is a standard joystick.

A non-proportional interface is one which involves a number of switches. Selecting a particular switch determines the direction of the wheelchair, but the speed is pre-programmed. One example of a non-proportional interface is a sip-and-puff mechanism.

The term controller describes the microprocessor and other related electronics that receive and interpret input from the joystick (or other drive control interface) and convert that input into power output which controls speed and direction. A high power wire harness connects the controller to the motor and gears.

A non-expandable controller has the following features:

- May have the ability to control up to 2 power seating actuators through the drive control (for example, seat elevator and single actuator power elevating legrests). (Note: Control of the power seating actuators through the Control Input Device would require the use of an additional component, E2310 or E2311.)
- Can accommodate only an integral joystick or a standard proportional remote joystick

- May allow for the incorporation of an attendant control

An expandable controller is capable of accommodating one or more of the following additional functions:

- Other types of proportional input devices (e.g., mini-proportional or compact joysticks, touchpads, chin control, head control, etc.)
- Non-proportional input devices (e.g., sip and puff, head array, etc.)
- Operate 3 or more powered seating actuators through the drive control. (Note: Control of the power seating actuators through the Control Input Device would require the use of an additional component, E2310 or E2311.)

An expandable controller may also be able to operate one or more of the following:

- A separate display (i.e., for alternate control devices)
- Other electronic devices (e.g., control of an augmentative speech device or computer through the chair's drive control)
- An attendant control

For power wheelchairs which are capable of being upgraded to an expandable controller (K0835 -K0891), E2377 is used if an expandable controller is provided at the time of initial issue. Code E2376 is used with complete replacement of an expandable controller.

A harness (E2313) describes all of the wires, fuse boxes, fuses, circuits, switches, etc. that are required for the operation of an expandable controller. It also includes all the necessary fasteners, connectors, and mounting hardware. Code E2313 is separately billable in addition to an expandable controller both at initial issue and with complete replacement of the expandable controller. Code K0108 must not be used for any component or feature of an expandable controller at the time of initial issue. The reimbursement for any type of complete expandable controller is included in the allowance for codes E2377/E2376 plus E2313. However, if individual components of the harness are replaced, code K0108 should be used.

A switch is an electronic device which turns power to a particular function either "on" or "off". The external component of a switch may be either mechanical or non-mechanical. Mechanical switches involve physical contact in order to be activated. Examples of the external components of mechanical switches include, but are not limited to, toggle, button, ribbon, etc. Examples of the external components of non-mechanical switches include, but are not limited to, proximity, infrared, etc. Some of the codes include multiple switches. In those situations, each functional switch may have its own external component or multiple functional switches may be integrated into a single external switch component or multiple functional switches may be integrated into the wheelchair control interface without having a distinct external switch component.

A stop switch allows for an emergency stop when a wheelchair with a non-proportional interface is operating in the latched mode. (Latched mode is when the wheelchair continues to move without the beneficiary having to continually activate the interface.) This switch is sometimes referred to as a kill switch.

A direction change switch allows the beneficiary to change the direction that is controlled by another separate switch or by a mechanical proportional head control interface. For example, it allows a switch to initiate forward movement one time and backward movement another time.

A function selection switch allows the beneficiary to determine what operation is being controlled by the interface at any particular time. Operations may include, but are not limited to, drive forward, drive backward, tilt forward, recline backward, etc.

An integrated proportional joystick and controller is an electronics package in which a joystick and controller electronics are in a single box, which is mounted on the arm of the wheelchair.

The interfaces described by codes E2312, E2321, E2322, E2325, E2327-E2330, and E2373-E2377 must have programmable control parameters for speed adjustment, tremor dampening, acceleration control, and braking.

A remote joystick is one in which the joystick is in one box that is typically mounted on the arm of the wheelchair and the controller electronics are located in a different box that is typically located under the seat of the wheelchair. The joystick is connected to the controller through a low power wire harness. A remote joystick may be used for either hand control, chin control, or attendant control.

A standard proportional remote joystick is one which requires approximately 340 grams of force to activate and which has an excursion (length of throw) of approximately 25 mm from neutral position. It can be used with a non-expandable or an expandable controller. There is no separate billing for a standard proportional remote joystick when it is provided at the time of initial issue of a power wheelchair whether it is used for hand or chin control by the beneficiary or whether it is used as an attendant control in place of a beneficiary-operated drive control interface.

A mini-proportional (short throw) remote joystick (E2312) is one which can be activated by a very low force (approximately 25 grams) and which has a very short displacement (a maximum excursion of approximately 5 mm from neutral). It can only be used with an expandable controller. It can be used for hand or chin control or control by other body part (e.g., tongue, lip, fingertip, etc.). There is no separate billing for control buttons, displays, switches, etc. There is no separate billing for fixed mounting hardware, regardless of the body part used to activate the joystick.

A compact proportional remote joystick (E2373) is one which has a maximum excursion of about 15 mm from neutral position but requires approximately 340 grams of force to activate. It can only be used with an expandable controller. It can be used for hand or chin control or control by other body part (e.g., foot, amputee stump, etc.). There is no separate billing for control buttons, displays, switches, etc. There is no separate billing for fixed mounting hardware, regardless of the body part used to activate the joystick.

A touchpad is an interface similar to the pad-type mouse found on portable computers. It is billed with code K0108.

Code E2321 is used for a non-proportional remote joystick, regardless of whether it is used for hand or chin control.

When code E2312, E2321, E2373, or E2374 is used for a chin control interface, the chin cup is billed separately with code E2324.

Code E2322 describes a system of 3-5 mechanical switches which are activated by the beneficiary touching the switch. The switch that is selected determines the direction of the wheelchair. A mechanical stop switch and a mechanical direction change switch, if provided, are included in the allowance for the code.

Code E2323 includes prefabricated joystick handles that have shapes other than a straight stick - e.g., U shape or T shape - or that have some other nonstandard feature - e.g., flexible shaft.

A sip and puff interface (E2325) is a non-proportional interface in which the beneficiary holds a tube in their mouth and controls the wheelchair by either sucking in (sip) or blowing out (puff). A mechanical stop switch is included in the allowance for the code. E2325 does not include the breath tube kit which is described by code E2326.

A proportional, mechanical head control interface (E2327) is one in which a headrest is attached to a joystick-like device. The direction and amount of movement of the beneficiary's head pressing on the headrest control the direction and speed of the wheelchair. A mechanical direction control switch is included in the code.

A proportional, electronic head control interface (E2328) is one in which a beneficiary's head movements are sensed by a box placed behind the beneficiary's head. The direction and amount of movement of the beneficiary's head (which does not come in contact with the box) control the direction and speed of the wheelchair. A proportional, electronic extremity control interface (E2328) is one in which the direction and amount of movement of the beneficiary's arm or leg control the direction and speed of the wheelchair.

A non-proportional, contact switch head control interface (E2329) is one in which a beneficiary activates one of three mechanical switches placed around the back and sides of their head. These switches are activated by pressure of the head against the switch. The switch that is selected determines the direction of the wheelchair. A mechanical stop switch and a mechanical direction change switch are included in the allowance for the code.

A non-proportional, proximity switch head control interface (E2330) is one in which a beneficiary activates one of three switches placed around the back and sides of their head. These switches are activated by movement of the

head toward the switch, though the head does not touch the switch. The switch that is selected determines the direction of the wheelchair. A mechanical stop switch and a mechanical direction change switch is included in the allowance for the code.

An attendant control is one which allows a caregiver to drive the wheelchair instead of the beneficiary. The attendant control is usually mounted on one of the rear canes of the wheelchair. This code is limited to proportional control devices, usually a joystick. Code E2331 is used when an attendant control is provided in addition to a beneficiary-operated drive control interface.

Codes E2374-E2376 describe components of drive control systems. They may only be used for replacements other than at the time of initial issue.

Code K0108 is appropriately used at the time of initial issue only when the drive control interface that is provided is not included in the base code and there is no specific E code which describes it. K0108 must not be used for additional features of a joystick.

Code K0108 is appropriately used at the time of replacement in the following situations:

1. An integrated proportional joystick and controller box are being replaced due to damage; or
2. An interface other than a remote joystick (e.g. sip and puff, head control) is being replaced but the controller is not being replaced; or
3. There is no specific E code which describes the type of drive control interface system which is provided.

The KC modifier (replacement of special power wheelchair interface) is used in the following situations:

1. Due to a change in the beneficiary's condition an integrated joystick and controller is being replaced by another drive control interface - e.g., remote joystick, head control, sip and puff, etc.; or
2. The beneficiary had a drive control interface described by codes E2321-E2322, E2325, E2327-E2330, or E2373 and both the interface (e.g., joystick, head control, sip and puff) and the controller electronics are being replaced due to irreparable damage.

The KC modifier would never be used at the time of initial issue of a wheelchair. The KC modifier specifically states replacement, therefore, the RB modifier is not required.

OTHER POWER WHEELCHAIR ACCESSORIES:

A drive wheel is one which is directly controlled by the motor of the power wheelchair. It may be either a rear wheel, mid wheel, or front wheel, depending on the model of the power wheelchair.

A caster is a smaller wheel that is in contact with the ground during normal operation of the wheelchair and which not directly controlled by the motor. It may be in the front and/or rear, depending on the location of the drive wheel.

A pneumatic tire (E2381, E2384) is a rubber tire which is used in conjunction with a separate tube (E2382, E2385) which is filled with air.

A flat free insert (E2383) is a removable ring of firm material that is placed inside of a pneumatic tire to allow the wheelchair to continue to move if the pneumatic tire is punctured. This code may not be used for a foam filled tire.

A foam filled tire (E2386, E2387) is one in which a rubber tire shell has been filled with foam which is non-removable.

A foam tire (E2388, E2389) is one which is made entirely of self-skinning urethane.

A solid tire (E2390, E2391, E2392) is one which is made of hard plastic or rubber.

All types of tires and wheels are included in the code for a power mobility base. Codes E2381-E2396 may only be used for replacements other than at the time of initial issue.

Code E2351 describes an electronic interface used with a speech generating device. An electronic interface that is used to allow lights or other electrical devices to be operated using the power wheelchair control interface must be billed with code A9270 (non-covered item).

Codes E2368-E2370 are for a replacement motor and/or gearbox. These codes are not used at the time of initial issue. If the item is a rebuilt component, the UE (used DME) modifier must be added to the code.

MISCELLANEOUS:

Code E1028 is used for

1. Swingaway hardware used with remote joysticks or touchpads,
2. Swingaway or flip-down hardware for head control interfaces E2327-E2330, and
3. Swingaway hardware for an indicator display box that is related to the multi-motor electronic connection codes E2310 or E2311.

Code E1028 is not to be used for swingaway hardware used with a sip and puff interface (E2325) because swingaway hardware is included in the allowance for that code. Code E1028 is not to be used for hardware on a wheelchair tray (E0950). Do not use E1028 in addition to E1020 (Residual limb support system) as it includes swingaway hardware.

Code E1029 describes a ventilator tray which is attached in a fixed position to the wheelchair base or back. Code E1030 describes a ventilator tray which is attached to the seat back and is articulated so that the tray will remain horizontal when the seat back is raised or lowered.

Code E1225 describes a manually operated reclining back that can recline greater than 15 degrees but less than 80 degrees. Code E1226 describes a manually operated reclining back that reclines 80 degrees or greater.

A Column II code is included in the allowance for the corresponding Column I code when provided at the same time. When multiple codes are listed in column I, all the codes in column II relate to each code in column I.

Column I	Column II
Power Operated Vehicle (K0800-K0812)	All options and accessories
Rollabout Chair (E1031)	All options and accessories
Transport Chair (E1037, E1038, E1039)	All options and accessories except E0990, K0195
Manual Wheelchair Base (E1161, E1229, E1231, E1232, E1233, E1234, E1235, E1236, E1237, E1238, K0001, K0002, K0003, K0004, K0005, K0006, K0007, K0009)	E0967, E0981, E0982, E0995, E2205, E2206, E2210, E2220, E2221, E2222, E2224, E2225, E2226, K0015, K0017, K0018, K0019, K0042, K0043, K0044, K0045, K0046, K0047, K0050, K0052, K0069, K0070, K0071, K0072, K0077
Power Wheelchair Base Groups 1 and 2 (K0813-K0843)	E0971, E0978, E0981, E0982, E0995, E1225, E2366, E2367, E2368, E2369, E2370, E2374, E2375, E2376, E2378, E2381, E2382, E2383, E2384, E2385, E2386, E2387, E2388, E2389, E2390, E2391, E2392, E2394, E2395, E2396, K0015, K0017, K0018, K0019, K0037, K0040, K0041, K0042, K0043, K0044, K0045, K0046, K0047, K0051, K0052, K0077, K0098
Power Wheelchair Base Groups 3, 4, and 5 (K0848-K0891)	E0971, E0978, E0981, E0982, E0995, E1225, E2366, E2367, E2368, E2369, E2370, E2374, E2375, E2376, E2378, E2381, E2382, E2383, E2384, E2385, E2386, E2387, E2388, E2389, E2390, E2391, E2392, E2394, E2395, E2396, K0015, K0017, K0018, K0019, K0037, K0041, K0042, K0043, K0044, K0045, K0046, K0047, K0051, K0052, K0077, K0098
E0973	K0017, K0018, K0019
E0950	E1028
E0990	E0995, K0042, K0043, K0044, K0045, K0046, K0047
Power tilt and/or recline seating systems (E1002, E1003, E1004, E1005, E1006, E1007, E1008)	E0973, K0015, K0017, K0018, K0019, K0020, K0042, K0043, K0044, K0045, K0046, K0047, K0050, K0051, K0052
E1009, E1010, E1012	

	E0990, E0995, K0042, K0043, K0044, K0045, K0046, K0047, K0052, K0053, K0195
E2325	E1028
E1020	E1028
K0039	K0038
K0045	K0043, K0044
K0046	K0043
K0047	K0044
K0053	E0990, E0995, K0042, K0043, K0044, K0045, K0046, K0047
K0069	E2220, E2224
K0070	E2211, E2212, E2224
K0071	E2214, E2215, E2225, E2226
K0072	E2219, E2225, E2226
K0077	E2221, E2222, E2225, E2226
K0195	E0995, K0042, K0043, K0044, K0045, K0046, K0047

[Back to Top](#)

[Coding Information](#)

Bill Type Codes:

Contractors may specify Bill Types to help providers identify those Bill Types typically used to report this service. Absence of a Bill Type does not guarantee that the article does not apply to that Bill Type. Complete absence of all Bill Types indicates that coverage is not influenced by Bill Type and the article should be assumed to apply equally to all claims.

N/A

Revenue Codes:

Contractors may specify Revenue Codes to help providers identify those Revenue Codes typically used to report this service. In most instances Revenue Codes are purely advisory. Unless specified in the article, services reported under other Revenue Codes are equally subject to this coverage determination. Complete absence of all Revenue Codes indicates that coverage is not influenced by Revenue Code and the article should be assumed to apply equally to all Revenue Codes.

N/A

CPT/HCPCS Codes N/A

ICD-10 Codes that are Covered N/A

ICD-10 Codes that are Not Covered N/A

[Back to Top](#)

[Revision History Information](#)

Revision History Date	Revision History Number	Revision History Explanation
01/01/2017	R5	Revision Effective Date: 01/01/2017 NON-MEDICAL NECESSITY COVERAGE AND PAYMENT RULES:

Revision History Date	Revision History Number	Revision History Explanation
01/01/2017	R4	<p>Removed: Non-coverage language for E0950 due to clerical error Revision Effective Date: 01/01/2017 NON MEDICAL NECESSITY PAYMENT RULES: Added: Coverage rule for E0950 POLICY SPECIFIC DOCUMENTATION REQUIREMENTS: Added: 42 CFR 410.38(g) and Modifier requirements Added: Instructions for replacement codes RELATED LOCAL COVERAGE DOCUMENTS: Added: LCD-related Standard Documentation Requirements article</p>
07/01/2016	R3	<p>Effective July 1, 2016 oversight for DME MAC Articles is the responsibility of CGS Administrators, LLC 18003 and 17013 and Noridian Healthcare Solutions, LLC 19003 and 16013. No other changes have been made to the Articles.</p>
01/01/2016	R2	<p>Revision Effective Date: 01/01/2016 NON-MEDICAL NECESSITY COVERAGE AND PAYMENT RULES: Revised: Standard Documentation Language to remove start date verbiage from Prescription Requirements (Effective 11/5/2015) CODING GUIDELINES: Added: HCPCS code E1012 Added: HCPCS code E1012 to bundling table</p>
10/01/2015	R1	<p>Revision Effective Date: 10/31/2014 NON-MEDICAL NECESSITY COVERAGE AND PAYMENT RULES: Removed: "When required by state law" from ACA new prescription requirements Revised: Face-to-Face Requirements for treating practitioner CODING GUIDELINES: Revised: Removed HCPCS K0017 and K0018 from the initial package verbiage for armrest separate billing due to being parts of the whole assembly E0973 and only separately billed for replacement parts. Added: E0973 was added to the initial package verbiage for armrest separate billing due to being the whole assembly. Removed: The word "adjustable" was removed from the initial package verbiage for armrest separate billing due to fixed armrests K0020 being included</p>

[Back to Top](#) **Related Local Coverage Document(s)** Article(s) [A55426 - Standard Documentation Requirements for All Claims Submitted to DME MACs](#) LCD(s) [L33792 - Wheelchair Options/Accessories](#)

Related National Coverage Document(s) N/A

Statutory Requirements URL(s) N/A

Rules and Regulations URL(s) N/A

CMS Manual Explanations URL(s) N/A

Other URL(s) N/A

Public Version(s) Updated on 03/24/2017 with effective dates 01/01/2017 - N/A [Updated on 03/09/2017 with effective dates 01/01/2017 - N/A](#) [Updated on 06/07/2016 with effective dates 07/01/2016 - N/A](#) [Updated on 02/26/2016 with effective dates 01/01/2016 - N/A](#) Some older versions have been archived. Please visit [MCD Archive Site](#) to retrieve them. [Back to Top](#)

[Keywords](#)

N/A Read the [Article Disclaimer](#) [Back to Top](#)